

-----  
Mauna Loa Solar Observatory Observer's Log  
-----

Mon Sep 23 16:19:36 GMT 2013

Year: 13 Doy: 266

Observer: berkey

WEATHER COMMENT: berkey: Mon Sep 23 16:19:38 GMT 2013

clear skies, temp 45f, wind 8mph from SE

\_\_\_\_end\_\_\_\_

Mon Sep 23 16:38:57 GMT 2013: PSPT Start Patrol

Mon Sep 23 16:42:55 GMT 2013 COMP Start Patrol

GENERAL OBSERVATORY COMMENT BY berkey: Mon Sep 23 19:24:52 GMT 2013

Cirrus showed up

\_\_\_\_end\_\_\_\_

Mon Sep 23 19:28:06 GMT 2013 COMP End Patrol

Mon Sep 23 20:47:38 GMT 2013 COMP Start Patrol

Mon Sep 23 22:07:06 GMT 2013: PSPT Start Patrol

GENERAL OBSERVATORY COMMENT BY berkey: Mon Sep 23 22:47:30 GMT 2013

cirrus has come back.

\_\_\_\_end\_\_\_\_

Mon Sep 23 22:50:06 GMT 2013 COMP End Patrol

Mon Sep 23 23:17:20 GMT 2013 COMP Start Patrol

Mon Sep 23 23:51:01 GMT 2013 COMP End Patrol

Mon Sep 23 23:58:29 GMT 2013 COMP Start Patrol

Tue Sep 24 00:07:26 GMT 2013 COMP End Patrol

Tue Sep 24 01:59:12 GMT 2013 COMP Start Patrol

GENERAL OBSERVATORY COMMENT BY berkey: Tue Sep 24 01:59:57 GMT 2013

Architects came up to inspect the building as a pre-bid for future work.

\_\_\_\_end\_\_\_\_

Tue Sep 24 02:10:59 GMT 2013 COMP End Patrol

Tue Sep 24 03:24:42 GMT 2013: PSPT Abort Patrol

Tue Sep 24 03:24:46 GMT 2013: PSPT Abort Patrol

KCOR COMMENT BY berkey: Tue Sep 24 03:33:32 GMT 2013

Today I through together an observing version of the k-cor observing code designed around goal of starting socket came on (or maybe twice) a day and letting it run continuously. This means that we will be getting a lot of bad/garbage bin and fits frames any time an optic is moving or we are otherwise doing funky things to the instrument.

But I am 99% sure that all of the funky/moving things should be tagged as Fits datatype=engineering.

I re-used Brandon's calibration script parser. But now instead of starting socketcam to take an image; I just wait a few seconds for things to settle out then once I am sure we have a valid calibration frame I change the Fits header for the next image to calibration (from engineering) for 48 seconds. I think right now it is nice to have more than once calibration image at each state so we can make sure things are consistent. Soon we can dial the number back a bit so we only record one image at each state.

After calibration imaging is done I change the datatype back to engineering and proceed to move the stage.

The calibration script should make a note of what ini file it ran in this log, and create a log of all the commands with start/stop images for each state in e:\date\avg\calibration.txt

There is a observer trigger button that turns on datatype=science in the fits header. When this is triggered on/off a start/stop synoptic comment will appear in this log.

A first draft of a k-cor operation manual appears on the desktop of the k-cor machine next to the two VI's that are important for running it.

\_\_\_end\_\_\_