```
Mauna Loa Solar Observatory Observer's Log
______
      Thu Oct 10 17:05:44 GMT 2013
Year: 13 Doy: 283
Observer: stueben
WEATHER COMMENT: stueben: Thu Oct 10 17:05:53 GMT 2013
Overcast at present, wind east 5-10 mph, temp 45F. Awaiting clearer skies before
opening domes.
end
Thu Oct 10 19:19:08 GMT 2013: PSPT Start Patrol
WEATHER COMMENT: stueben: Thu Oct 10 19:19:56 GMT 2013
Clouds have thinned, pspt started.
end
Thu Oct 10 19:29:31 GMT 2013: PSPT Abort Patrol
Thu Oct 10 19:29:35 GMT 2013: PSPT Abort Patrol
Thu Oct 10 19:29:39 GMT 2013: PSPT Abort Patrol
Thu Oct 10 19:29:43 GMT 2013: PSPT Abort Patrol
Thu Oct 10 19:31:20 GMT 2013: PSPT Start Patrol
Thu Oct 10 21:40:16 GMT 2013: PSPT Abort Patrol
WEATHER COMMENT: stueben: Thu Oct 10 22:08:16 GMT 2013
In orographic clouds, instruments idled.
end
KCOR COMMENT BY stueben: Fri Oct 11 00:52:54 GMT 2013
Both filter timer lights were on for the lower HEPA unit. The timers had never
been configured and were in test mode. Reset the timers and set both Hepa units
for the recommended timer periods - 60 days for carbon filter and 2 yrs for the
HEPA filter.
```

____end___