Mauna Loa Solar Observatory Observer's Log Sat Dec 11 17:34:48 GMT 2021 Year: 21 Doy: 345 Observer: mlso WEATHER COMMENT: mcotter: Sat Dec 11 17:36:49 GMT 2021 Temp: 44.5f, Humidity: 30%, Pressure: 28.746in, Wind: 13mph from 163degs, Skies: Beauti ful clear skies with moderate winds from the south. The inversion layer is not discerni ble on the horizon. _end_ Sat Dec 11 17:53:44 GMT 2021 Kcor Focus/alignment program exited Sat Dec 11 17:58:00 GMT 2021 Kcor Focus/alignment program exited KCOR COMMENT BY mcotter: Sat Dec 11 18:01:44 GMT 2021 The Kcor instrument in now running, though the sky is a little bright. SGS offsets: X(RA): 50, Y(Dec): 5. Polarization checked good: Mid, Bright, Dark, Mid. end GENERAL COMMENT BY mcotter: Sat Dec 11 18:01:55 GMT 2021 Opened windows upstairs end GENERAL COMMENT BY mcotter: Sat Dec 11 18:02:06 GMT 2021 PM Blew off UCoMP 01 end GENERAL COMMENT BY mcotter: Sat Dec 11 18:02:14 GMT 2021 PM Blew off Kcor O1 end UCOMP COMMENT BY mcotter: Sat Dec 11 18:14:54 GMT 2021 I started Ucomp as normal but when I put it in "Center Occulter" the program just hangs up, the Ucomp mech-controller does nothing and the diffuser and the shutter are stayi ng in beam. I am going to try and exit the program and restart to see if it make a diff erence. end **KCOR PROBLEM COMMENT BY mcotter** : Sat Dec 11 18:23:18 GMT 2021 After trying to restart the Ucomp Controller program I again tried running the program but again it is hanging up. I am now power cycling the computer to see if it will have a positive effect. end WEATHER COMMENT: mcotter: Sat Dec 11 18:56:53 GMT 2021 High altitude Cirrus clouds are visible above the observatory and are stretching out ac ross the sky to the south and west. end KCOR COMMENT BY mcotter: Sat Dec 11 18:57:47 GMT 2021 The Kcor instrument has been paused due to bright skies and high altitude Cirrus in the viewing area. _end_ **UCoMP PROBLEM COMMENT BY mcotter** : Sat Dec 11 20:45:51 GMT 2021 I have power cycled the Ucomp computer multiple times but the instrument program contin ues to get hung up when I try to run it. I will continue to try and trouble shoot the p rogram, but I would not be able to run the instrument anyway ,as the sky is too bright and has high altitude Cirrus clouds in the viewing area. end Sat Dec 11 21:23:34 GMT 2021 Running UCOMP Cookbook dark_80ms_2beam_16sums_BOTH.cbk lin e 0 Sat Dec 11 21:24:51 GMT 2021 Running UCOMP Cookbook 637_Scan.cbk line 0 Sat Dec 11 21:27:01 GMT 2021 Running UCOMP Cookbook 656_Scan.cbk line 0 Sat Dec 11 21:28:51 GMT 2021 Running UCOMP Cookbook 789_Scan.cbk line 0 Sat Dec 11 21:30:40 GMT 2021 Running UCOMP Cookbook 1074_Scan.cbk line 0 Sat Dec 11 21:31:52 GMT 2021 UCoMP Paused for clouds Sat Dec 11 21:32:59 GMT 2021 UCoMP Restarted from pause Sat Dec 11 21:33:31 GMT 2021 Running UCOMP Cookbook 1079_Scan.cbk line 0 Sat Dec 11 21:35:21 GMT 2021 Running UCOMP Cookbook all_wavelenght_coronal_flat.cbk lin e 0 Sat Dec 11 21:53:27 GMT 2021 Running UCOMP Cookbook all_wavelenght_coronal.cbk line 0 **UCoMP PROBLEM COMMENT BY mcotter** : Sat Dec 11 21:54:48 GMT 2021 After repeated attempts to get the instrument software running without hanging up I cal

led Ben to help me to troubleshoot the issue. When looking in the software logic block

diagram Ben saw that the "Filter Wheel" was not communicating from the instrument softw are, therefore when the software reached a point in the initialization of the instrumen t it was expecting to see an output from the Filter Wheel, and it was not there so the instrument just hung up waiting for a response. Ben asked me to physically inspect the Filter Wheel. I removed the cover and checked the cable going to the Filter Wheel and i t looked fine. I pulled the plug out of the Filter Wheel (printer style USB connector) and plugged it back in and Ben told me to wait as it appeared the communication was res tored. After a minute or so Ben confirmed that the Filter Wheel was again communicating . I noted to Ben that the female side of the connector felt like it had some play in it and when reconnecting the connector it was not a nice tight snap in fit. I put the cov er back on and came back to the control room where at this time Ben instructed me to ma nually power cycle the Ucomp cameras. This is accomplished by unplugging the power cord to the cameras via a small two receptacle power cable in the back of the Ucomp rack. Having done this the instrument was now communicating properly with the software. Ucomp now appears to be operating properly.

____end____

UCOMP COMMENT BY mcotter: Sat Dec 11 22:10:47 GMT 2021 It appears that it is necessary to power cycle the cameras as described in the previous entry when the Ucomp instrument software has been restarted. Also, if further issues p ersist with the Ucomp instrument software hanging up it may be necessary to unplug and then again plug in the USB cable that goes to the Filter Wheel. As described in my prev ious entry, I accomplished this by removing the cover and unplugging the connector at t he housing of the filter wheel. If we need to unplug and plug in the connector again f or a similar problem we can try unplugging the Filter Wheel connector at the bottom of the instrument, on the exterior bulkhead. The connector for the Filter Wheel is all the way at the back of the instrument adjacent to the south side bulkhead. Additionally, t he area of the Filter Wheel where the USB cable plugs in is right up against the interi or bulkhead of the instrument. Ben asked me to look online and see if there is a 90 deg connector available that will offer more room for the USB connector/cable to be insert ed into the Filter Wheel housing. When convenient I am going to remove the Ucomp cover and take photos of the Filter Wheel USB connection area and receptacle, so that we can see if an alternate 90 deg connector/cable is available that will properly fit the conf iguration. _end_

KCOR COMMENT BY mcotter: Sat Dec 11 22:11:27 GMT 2021 Though the sky is a little bright the Kcor instrument is running. end UCOMP COMMENT BY mcotter: Sat Dec 11 22:12:01 GMT 2021 The Ucomp instrument appears to be operating normally. _end_ Sat Dec 11 22:20:27 GMT 2021 UCoMP Paused for clouds Sat Dec 11 22:31:20 GMT 2021 UCoMP Restarted from pause GENERAL COMMENT BY mcotter: Sat Dec 11 22:33:37 GMT 2021 I paused both instruments for about ten minutes due to Cirrus clouds passing through th e viewing area, but we are now back on sky. The sky is a bit bright and aerosols are vi sible but the images look acceptable. ___end_ Sat Dec 11 22:52:19 GMT 2021 Running UCOMP Cookbook all_wavelenght_coronal.cbk line 23 Sat Dec 11 23:19:30 GMT 2021 Running UCOMP Cookbook dark_80ms_2beam_16sums_BOTH.cbk lin e 0 Sat Dec 11 23:20:47 GMT 2021 Running UCOMP Cookbook 637_Pol_Calibrate.cbk line 0 Sat Dec 11 23:24:52 GMT 2021 Running UCOMP Cookbook 656_Pol_Calibrate.cbk line 0 Sat Dec 11 23:28:56 GMT 2021 Running UCOMP Cookbook 789_Pol_Calibrate.cbk line 0 Sat Dec 11 23:32:59 GMT 2021 Running UCOMP Cookbook 1074_Pol_Calibrate.cbk line 0 Sat Dec 11 23:37:03 GMT 2021 Running UCOMP Cookbook 1079_Pol_Calibrate.cbk line 0 Sat Dec 11 23:41:07 GMT 2021 Running UCOMP Cookbook waves_1074_1hour.cbk line 0 Sun Dec 12 00:09:09 GMT 2021 UCoMP Paused for clouds GENERAL COMMENT BY mcotter: Sun Dec 12 01:29:30 GMT 2021 The sky became quite bright and both instrument were stopped at approximately 2:10pm HS Τ. _end_ Sun Dec 12 01:44:23 GMT 2021 UCoMP Restarted from pause UCOMP COMMENT BY mcotter: Sun Dec 12 01:50:52 GMT 2021 Took photos of the Ucomp Filter Wheel USB connector that was giving us problems this mo rning, so that we can find an alternate cable with a 90 deg connector head. end

GENERAL COMMENT BY mcotter: Sun Dec 12 01:51:22 GMT 2021

Emptied trash receptacles in control room. ____end___ GENERAL COMMENT BY mcotter: Sun Dec 12 01:54:19 GMT 2021 The day started out nice and clear but by mid morning high altitude Cirrus clouds moved into the viewing area from the south-southwest. After a short pause the sky cleared en ough to again run the instruments, though the sky was a bit bright. By approximately 2: 00pm HST the cirrus returned and the sky became entirely too bright to observe. ____end___ ONSITE STAFF: mcotter